



**The Commonwealth of Massachusetts**  
 Executive Office of Health and Human Services  
 Department of Public Health  
 Office of Emergency Medical Services  
 2 Boylston Street, 3<sup>rd</sup> Floor  
 Boston, MA 02116  
 617-753-7300  
 FAX: 617-753-7320

**DEVAL L. PATRICK**  
 GOVERNOR

**TIMOTHY P. MURRAY**  
 LIEUTENANT GOVERNOR

**JUDYANN BIGBY, MD**  
 SECRETARY

**JOHN AUERBACH**  
 COMMISSIONER

**TABLE OF APPROVED DEVICES**

The following devices and categories of devices have been reviewed by OEMS. Advisories are on the OEMS web page regarding usage and restrictions, if any. As with any device operated under the license of the ambulance service, the ambulance service must ensure that the specific device is FDA approved and that their EMTs are properly trained, refreshed in their training, and have demonstrated competency under the direction of their affiliate hospital medical director, as per the manufacturer's recommendations and usage, prior to allowing such device on the ambulance. **This list will be updated as new devices are approved by the Department.**

<b>TABLE OF DEVICES</b>			
<b>DEVICE</b>	<b>DESCRIPTION</b>	<b>APPROVAL DATE</b>	<b>RESTRICTIONS</b>
Autopulse™ or equivalent	Battery powered mechanical chest compression device	September 12, 2008	
King LT or equivalent	Supraglottic airway device used as an alternative to tracheal intubation or mask ventilation	May 2008	May be used with approval of Affiliate Hospital Medical Director.
Bougie	Rigid ETT plastic coated introducer placed blindly into the trachea to introduce the ETT over it.	~ 2005	May be used with approval of Affiliate Hospital Medical Director.
Bone Injection Gun or equivalent	An automatic intraosseous (IO) Spring Loaded infusion device	February 11, 2008	
Optical Laryngoscope or equivalent	Anatomically shaped laryngoscope with two separate channels: The optical channel: contains a high definition optical system. The guiding channel: holds the endotracheal tube (ETT) and guides it through the vocal cords	February 11, 2008	
CPR Vest	Pneumatically powered mechanical chest compression system using air tanks.	November 21, 2007	
GlideScope® or equivalent	Laryngoscope which uses optics to visualize a patient's airway.	November 21, 2007	

**TABLE OF DEVICES**

<b>DEVICE</b>	<b>DESCRIPTION</b>	<b>APPROVAL DATE</b>	<b>RESTRICTIONS</b>
Hemostatic Dressings	Hemostatic Dressings Only (not powders)	November 21, 2007	See Department letter addressing EMS Role
ADD-Vantage® System	Diltiazem HCl	July 17, 2007	See Department letter addressing EMS Role
Impedance Threshold Device	Adjunct device, which attaches to the BVM, and increases blood flow to the heart and brain during CPR.	September 6, 2006	
CO Co-Oximeter	Carbon Monoxide Detection unit	September 6, 2006	
IV Power Injector	IV Bolus of Medications	September 6, 2006	
Pedi- IO Device	Pedi- Drill Type Only	September 6, 2006	
Ventricular Assist Device (VAD); (LVAD), (RVAD)	ALL	February 10, 2005	See Department letter addressing EMS Transport
Peripherally Inserted Central Catheter (PICC) Lines	All	October 10, 2003	See Department letter addressing EMS Role
Adult Intraosseous (IO) Infusion Devices	Adult- operator manipulation for its penetration power, no pressurized or “gun-type” devices	October 10, 2003	See Department letter addressing EMS Role